



Czech Metrology Institute

Notified Body No. 1383, Okružní 31
638 00 Brno

EC-TYPE EXAMINATION CERTIFICATE

Number: TCM 142/10 – 4794

Addition 1

This addition replaces all previous versions of this certificate in full wording.

Issued by: **Český metrologický institut**
Okružní 31
638 00 Brno
Czech Republic

Notified Body No. 1383

In accordance with: point 3 of annex 2 to Government Order No. 464/2005 Coll. (annex B of the Directive 2004/22/EC) from 19 October 2005 that lays down technical requirements on measuring instruments and implements in Czech Republic Directive 2004/22/EC of the European Parliament and of the Council.

Manufacturer:
(Applicant) **MADDALENA S.p.A.**
Via G. B. Maddalena 2/4
33040 Povoletto (UD)
Italy

In respect of: **water meter – single jet, dry dial**
type: CD SD PLUS
Accuracy class: 2
Temperature class: T50 or T30/90

Valid until: **17 November 2020**

Document number: **0115-CS-A039-10**

Description: Essential characteristics, approved conditions and special conditions, if any, are described in this certificate. This certificate contains 7 pages.

Date of issue: 15 April 2011




.....
RNDr. Pavel Klenovský

Notified Body No.1383



CERTYFIKAT ZGODNOŚCI Nr CE/13/58/10/BE

System certyfikacji-1a (PKN-ISO/IEC Przewodnik 67)

Nazwa i adres posiadacza certyfikatu: **FERRO S.A.**
ul. Przemysłowa 7, 32-050 Skawina, Polska.

Nazwa i adres producenta: **FERRO S.A.**
ul. Przemysłowa 7, 32-050 Skawina, Polska.

Nazwa wyrobu: Pompa cyrkulacyjna

Typ (odmiany): CP-15-1.5

Znak handlowy WEBERMAN

Podstawowe parametry: 230V ~; 50-60Hz; 28W; klasa I

Wyrób spełnia wymagania zawarte w: EN 60335-1:2002+A11:2004+A1:2004+A12:2006+A2:2006+A13:2008;
EN 60335-2-51:2003+A1:2008; EN 50366:2003+A1:2006;
EN 55014-1:2006+A1:2009; EN 55014-2:1997+A1:2001+A2:2008;
EN 61000-3-2:2006+A1:2009+A2:2009; EN 61000-3-3:2008

Zgodnie ze sprawozdaniem z badań wykonanych przez: PCBC S.A.

Nr i data sprawozdania: BE/128/2010 z dnia 11.05.2010 r.;
BE/93/2010 z dnia 31.03.2010 r.

Wyniki badań potwierdzają spełnianie przez badany egzemplarz wyrobu wymagań norm zharmonizowanych z dyrektywą LVD 2006/95/WE, a analiza konstrukcji wyrobu potwierdza spełnianie wymagań zasadniczych zawartych w dyrektywie EMC 2004/108/WE. Po spełnieniu pozostałych wymagań zawartych w dyrektywach, w szczególności przygotowaniu wymaganej dokumentacji i wystawieniu deklaracji zgodności, producent lub jego upoważniony przedstawiciel może na własną odpowiedzialność umieścić oznakowanie CE na wyrobie zgodnym z wyrobem badanym.

DYREKTOR ds. BADAŃ i CERTYFIKACJI

Tadeusz Glazer



CERTIFICATE OF CONFORMITY No CE/13/7/10/BE

Product certification system-1a (ISO/IEC Guide 67).

Name and address of the certificate holder: **FERRO S.A.**
ul. Przemysłowa 7, 32-050 Skawina, Poland.

Name and address of the manufacturer: **FERRO S.A.**
ul. Przemysłowa 7, 32-050 Skawina, Poland.

Product: Circulation pump

Model/type Ref.: 1) 25-40-180; 2) 25-60-180

Trademark: WEBERMAN

Ratings: 230V ~; 50-60Hz; 65W⁽¹⁾; 100W⁽²⁾; class I

The product complies with the safety requirements set in the following standards: EN 60335-1:2002+A11:2004+A1:2004+A12:2006+A2:2006+A13:2008;
EN 60335-2-51:2003+A1:2008; EN 50366:2003+A1:2006;
EN 55014-1:2006+A1:2009; EN 55014-2:1997+A1:2001+A2:2008;
EN 61000-3-2:2006+A1:2009+A2:2009; EN 61000-3-3:2008

As shown in the test report carried out by: PCBC S.A.

Ref. No. and date of the test report: BE/137/2008 – 1/U/2010 of 26.01.2010;
BE/138/2008 – 1/U/2010 of 26.01.2010;
BE/141/2008 of 27.05.2008; BE/142/2008 of 26.05.2008

The test results confirm that the tested sample of product fulfils the requirements of harmonized standards against directive LVD 2006/95/EC and analysis of the product construction confirms the fulfillment of the essential requirements included in the directive EMC 2004/108/EC . In case the rest of the appropriate requirements of directives is fulfilled, especially after establishing the technical documentation file as well issuing the EC Declaration of Conformity, the manufacturer or his EU representative can affix the CE marking to product, on his own responsibility.

DIRECTOR for TESTING and CERTIFICATION


Tadeusz Glazer

Warsaw, 29.04.2011

Electrical Product Certification Department